
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001826
9:37 21/06/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:07
STERILIZING 4:12
136.0 C 220kPa 4:12
136.2 C 221kPa 4:42
136.3 C 222kPa 5:12
136.4 C 223kPa 5:42
136.2 C 222kPa 6:12
136.2 C 219kPa 6:42
136.1 C 221kPa 7:12
136.3 C 223kPa 7:42
136.3 C 223kPa 7:42
VENTING 7:42
AIR DRYING 8:35
CYCLE COMPLETE 8:37

Digital Signature #
9941E2E6FCD73D62
