

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1613

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2022 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.1 F 0.0  
2:00 163.3 F 2.7  
4:00 178.5 F 6.8  
6:00 200.8 F 12.1  
8:00 223.9 F 7.6  
10:00 237.2 F 13.2  
12:00 236.8 F 11.0  
14:00 237.2 F 9.8  
16:00 250.3 F 16.6  
18:00 262.3 F 23.5  
20:00 272.7 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.6  
1:00 272.7 F 30.5  
2:00 272.5 F 30.0  
3:00 271.7 F 29.5  
4:00 272.1 F 30.2

Min 271.5 F 29.5  
Max 272.9 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
5:00 204.6 F 0.0  
10:00 220.0 F 0.0  
15:00 217.8 F 0.0  
20:00 204.4 F 0.0  
25:00 200.1 F 0.0  
Dry Cycle Aborted

FILLING: 2:15  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 28:42  
TOTAL CYCLE: 00:58:31

10/18/2022 10:08 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02