
Midmark M11 - v1.0.0
Total Cycles: 2136

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2022 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.2 F 0.0
2:00 202.8 F 3.9
4:00 200.1 F 8.7
6:00 220.0 F 8.3
8:00 230.7 F 9.0
10:00 240.3 F 10.4
12:00 241.7 F 12.0
14:00 240.3 F 10.5
16:00 253.4 F 17.2
18:00 264.5 F 23.8
20:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.8 F 29.3
2:00 271.8 F 28.5
3:00 272.3 F 29.2
4:00 272.5 F 29.0

Min 271.5 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 228.5 F 0.0
10:00 228.5 F 0.0
15:00 233.0 F 0.0
20:00 220.9 F 0.0
25:00 220.2 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:21
CYCLE COMPLETE

07/15/2022 11:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____