
Midmark M11 - v1.0.5
Total Cycles: 489

Name: RITTER2
Sterilizer ID: V2313228
10/27/2021 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 178.0 F 2.6
4:00 180.2 F 7.1
6:00 201.2 F 12.5
8:00 224.8 F 8.2
10:00 234.8 F 12.3
12:00 237.6 F 12.0
14:00 239.9 F 11.0
16:00 253.2 F 17.6
18:00 264.9 F 24.4
20:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 273.1 F 30.3
1:00 272.4 F 30.3
2:00 272.1 F 30.5
3:00 272.2 F 30.5
4:00 272.3 F 30.5

Min 271.9 F 30.2
Max 273.1 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 231.0 F 0.0
10:00 239.9 F 0.0
15:00 219.7 F 0.0
20:00 186.7 F 0.0
25:00 178.3 F 0.0
Dry Cycle Aborted

FILLING: 1:57
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:11
DRYING: 26:32
TOTAL CYCLE: 00:56:36

10/27/2021 10:40 AM
ACCEPTED: X
REJECTED:
Operator: U01