
Midmark M11 - v1.0.0
Total Cycles: 3490

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2023 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 194.7 F 4.3
4:00 205.5 F 9.3
6:00 220.7 F 6.9
8:00 233.3 F 9.8
10:00 235.9 F 8.6
12:00 246.0 F 13.5
14:00 244.9 F 12.6
16:00 258.1 F 19.6
18:00 269.9 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.5 F 28.7
2:00 271.5 F 28.0
3:00 272.7 F 28.7
4:00 271.7 F 28.0

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 228.9 F 0.0
10:00 225.8 F 0.0
15:00 225.2 F 0.0
20:00 224.7 F 0.0
25:00 225.0 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

07/05/2023 03:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____