
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000531
13:46 17/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:49
STERILIZING 3:59
136.0 C 220kPa 3:59
136.2 C 232kPa 4:29
135.8 C 229kPa 4:59
136.1 C 229kPa 5:29
136.4 C 233kPa 5:59
136.2 C 229kPa 6:29
136.4 C 232kPa 6:59
136.4 C 231kPa 7:29
136.4 C 231kPa 7:29
VENTING 7:29
AIR DRYING 8:21
CYCLE COMPLETE 23:21

Digital Signature #
0B9C6FE1CCE82350
