
Midmark M11 - v1.0.5
Total Cycles: 574

Name: RITTER2
Sterilizer ID: V2313228
11/24/2021 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 178.2 F 4.2
4:00 189.3 F 9.4
6:00 213.7 F 6.1
8:00 225.9 F 10.4
10:00 231.5 F 8.0
12:00 239.9 F 10.9
14:00 247.2 F 14.2
16:00 260.1 F 21.4
18:00 271.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.8
2:00 271.8 F 29.2
3:00 271.0 F 28.9
4:00 271.8 F 29.8

Min 271.0 F 28.7
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 228.9 F 0.0
10:00 236.8 F 0.0
15:00 217.4 F 0.0
20:00 215.0 F 0.0
25:00 218.5 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 18:38
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:34
CYCLE COMPLETE

11/24/2021 02:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____