
Midmark M11 - v1.0.0
Total Cycles: 1996

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2022 02:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.6 F 0.0
2:00 178.7 F 4.9
4:00 192.9 F 9.5
6:00 220.3 F 6.8
8:00 224.9 F 9.2
10:00 235.5 F 9.5
12:00 240.0 F 11.3
14:00 237.9 F 9.5
16:00 252.0 F 16.2
18:00 263.6 F 23.0
20:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.7 F 28.9
2:00 271.4 F 28.1
3:00 272.4 F 29.1
4:00 272.1 F 28.6

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 230.9 F 0.0
10:00 225.1 F 0.0
15:00 227.4 F 0.0
20:00 229.0 F 0.0
25:00 229.3 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

06/06/2022 03:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____