
Midmark M11 - v1.0.0
Total Cycles: 4630

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2024 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.8 F 0.0
2:00 169.2 F 3.2
4:00 181.6 F 6.8
6:00 200.2 F 10.7
8:00 216.3 F 4.4
10:00 224.7 F 7.8
12:00 238.8 F 12.0
14:00 235.5 F 8.6
16:00 243.7 F 12.0
18:00 244.8 F 12.4
20:00 255.8 F 18.1
22:00 265.8 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.4 F 28.9
2:00 272.3 F 28.5
3:00 271.4 F 28.0
4:00 271.3 F 28.3

Min 271.0 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 194.7 F 0.0
10:00 195.8 F 0.0
15:00 205.0 F 0.0
20:00 199.6 F 0.0
25:00 201.5 F 0.0
30:00 189.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 23:55
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

06/12/2024 12:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____