
Midmark M11 - v1.0.0
Total Cycles: 4118

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2023 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 157.2 F 2.4
4:00 193.7 F 6.1
6:00 214.8 F 11.4
8:00 221.4 F 4.9
10:00 233.7 F 9.4
12:00 238.8 F 10.2
14:00 240.7 F 11.0
16:00 237.2 F 8.9
18:00 250.7 F 15.3
20:00 262.3 F 21.9
22:00 272.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.5 F 28.4
2:00 271.6 F 28.3
3:00 272.3 F 28.3
4:00 271.5 F 28.3

Min 271.1 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 208.9 F 0.0
10:00 247.0 F 0.0
15:00 236.6 F 0.0
20:00 225.7 F 0.0
25:00 223.1 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

12/14/2023 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____