

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3990

Name: RITTER2  
Sterilizer ID: V2313228  
10/02/2024 02:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.8 F 0.0  
2:00 154.2 F 2.0  
4:00 184.0 F 5.5  
6:00 198.3 F 9.7  
8:00 219.7 F 9.0  
10:00 225.4 F 8.5  
12:00 233.6 F 13.0  
14:00 236.7 F 9.9  
16:00 237.7 F 9.5  
18:00 246.4 F 13.8  
20:00 257.4 F 19.7  
22:00 267.4 F 25.9  
24:00 273.1 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 271.7 F 29.2  
2:00 272.1 F 29.3  
3:00 272.6 F 29.5  
4:00 272.3 F 29.7

Min 271.2 F 29.2  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 247.9 F 0.0  
10:00 218.5 F 0.0  
15:00 185.7 F 0.0  
20:00 178.9 F 0.0  
25:00 177.3 F 0.0  
30:00 176.2 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 24:13  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:03:28  
CYCLE COMPLETE

10/02/2024 03:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_