
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004900
13:16 04/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 1:57
STERILIZING 2:55
136.4 C 220kPa 2:55
136.3 C 223kPa 3:25
136.4 C 222kPa 3:55
136.4 C 224kPa 4:25
136.4 C 225kPa 4:55
136.4 C 223kPa 5:25
136.3 C 222kPa 5:55
136.3 C 224kPa 6:25
136.3 C 224kPa 6:25
VENTING 6:25
AIR DRYING 7:13
CYCLE COMPLETE 14:14

Digital Signature #
29C5CA9B57D11B32
