

-----  
Midmark M11 - v1.0.5  
Total Cycles: 484

Name: RITTER2  
Sterilizer ID: V2313228  
10/26/2021 10:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.2 F 0.0  
2:00 140.9 F 2.1  
4:00 159.1 F 5.2  
6:00 181.1 F 9.6  
8:00 200.3 F 6.0  
10:00 230.0 F 8.5  
12:00 235.9 F 11.0  
14:00 241.9 F 12.4  
16:00 240.8 F 11.3  
18:00 254.1 F 18.0  
20:00 265.2 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 273.0 F 29.8  
2:00 271.6 F 29.6  
3:00 272.6 F 29.8  
4:00 271.5 F 29.4  
  
Min 271.5 F 29.2  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.8 F 0.0  
5:00 226.8 F 0.0  
10:00 221.6 F 0.0  
15:00 225.0 F 0.0  
20:00 199.8 F 0.0  
25:00 180.9 F 0.0  
30:00 190.3 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 21:49  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:13  
CYCLE COMPLETE

10/26/2021 11:59 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01