
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005743
7:43 13/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:19
STERILIZING 4:26
136.4 C 223kPa 4:26
136.1 C 221kPa 4:56
136.4 C 218kPa 5:26
136.4 C 222kPa 5:56
135.9 C 219kPa 6:26
136.1 C 222kPa 6:56
136.3 C 222kPa 7:26
135.8 C 220kPa 7:56
135.8 C 220kPa 7:56
VENTING 7:56
AIR DRYING 8:42
CYCLE COMPLETE 8:50

Digital Signature #
DE691B446BD82FAE
