
Midmark M11 - v1.0.0
Total Cycles: 3951

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2023 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 189.4 F 3.1
4:00 197.8 F 7.0
6:00 216.2 F 12.2
8:00 224.7 F 6.4
10:00 236.1 F 11.2
12:00 234.9 F 8.1
14:00 243.5 F 12.0
16:00 244.7 F 12.7
18:00 256.8 F 19.0
20:00 267.3 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.2 F 29.1
2:00 271.9 F 28.6
3:00 271.6 F 28.8
4:00 272.3 F 28.9

Min 271.2 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 214.5 F 0.0
10:00 222.2 F 0.0
15:00 227.6 F 0.0
20:00 225.5 F 0.0
25:00 222.3 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:00:42
CYCLE COMPLETE

10/27/2023 12:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____