

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 002027  
13:46 12/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:06  
PRESSURIZING 2:13  
STERILIZING 3:12  
136.5 C 220kPa 3:12  
136.3 C 225kPa 3:42  
136.2 C 223kPa 4:12  
136.4 C 224kPa 4:42  
136.3 C 225kPa 5:12  
136.4 C 226kPa 5:42  
136.2 C 223kPa 6:12  
136.4 C 226kPa 6:42  
136.4 C 226kPa 6:42  
VENTING 6:42  
AIR DRYING 7:38  
CYCLE COMPLETE 26:00  
-----

Digital Signature #  
264825590F76C8B0  
-----