
Midmark M11 - v1.0.0
Total Cycles: 4471

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2024 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.6 F 0.0
2:00 165.4 F 1.8
4:00 183.5 F 4.9
6:00 198.0 F 9.2
8:00 211.4 F 13.4
10:00 223.9 F 6.2
12:00 233.7 F 10.3
14:00 234.5 F 8.5
16:00 242.4 F 11.7
18:00 238.9 F 9.8
20:00 251.2 F 15.7
22:00 261.8 F 21.7
24:00 271.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.6
1:00 273.0 F 29.1
2:00 271.6 F 28.8
3:00 272.4 F 28.8
4:00 272.4 F 29.2

Min 271.4 F 28.5
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 199.7 F 0.0
10:00 212.9 F 0.0
15:00 198.8 F 0.0
20:00 197.2 F 0.0
25:00 191.1 F 0.0
30:00 188.5 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:24
CYCLE COMPLETE

04/18/2024 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____