
Midmark M11 - v1.0.5
Total Cycles: 2944

Name: RITTER2
Sterilizer ID: V2313228
09/28/2023 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.7 F 0.0
2:00 180.7 F 3.7
4:00 202.7 F 9.2
6:00 220.9 F 7.5
8:00 233.5 F 11.0
10:00 233.8 F 8.8
12:00 239.7 F 10.0
14:00 248.7 F 14.9
16:00 261.2 F 21.9
18:00 272.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.3 F 29.2
2:00 271.4 F 28.8
3:00 272.4 F 29.2
4:00 271.4 F 28.5

Min 271.3 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 214.4 F 0.0
10:00 221.4 F 0.0
15:00 220.4 F 0.0
20:00 223.4 F 0.0
25:00 226.2 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:27
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

09/28/2023 12:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____