
Midmark M11 - v1.0.0
Total Cycles: 2186

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 187.2 F 3.5
4:00 192.0 F 7.9
6:00 216.2 F 12.9
8:00 224.4 F 7.4
10:00 236.8 F 12.4
12:00 235.1 F 9.2
14:00 240.1 F 9.7
16:00 246.2 F 13.5
18:00 258.1 F 19.9
20:00 268.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.3
2:00 272.5 F 29.0
3:00 271.6 F 28.4
4:00 271.6 F 28.8

Min 271.3 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 233.4 F 0.0
10:00 227.2 F 0.0
15:00 230.3 F 0.0
20:00 228.9 F 0.0
25:00 231.5 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

07/27/2022 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____