

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5021

Name: RITTER 1  
Sterilizer ID: V2178198  
10/24/2024 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.4 F 0.0  
2:00 134.1 F 0.4  
4:00 159.5 F 0.9  
6:00 182.0 F 3.4  
8:00 203.4 F 7.3  
10:00 217.7 F 12.0  
12:00 223.6 F 6.2  
14:00 234.7 F 11.1  
16:00 232.4 F 7.7  
18:00 244.8 F 12.4  
20:00 244.5 F 12.4  
22:00 257.0 F 18.9  
24:00 268.2 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.6 F 28.8  
2:00 271.7 F 28.1  
3:00 272.0 F 28.6  
4:00 272.4 F 28.7

Min 271.3 F 27.9  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.8 F 0.0  
5:00 199.2 F 0.0  
10:00 196.2 F 0.0  
15:00 214.8 F 0.0  
20:00 227.0 F 0.0  
25:00 230.7 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 25:19  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:03:35  
CYCLE COMPLETE

10/24/2024 01:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_