
Midmark M11 - v1.0.5
Total Cycles: 2769

Name: RITTER2
Sterilizer ID: V2313228
08/15/2023 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.0 F 0.0
2:00 181.7 F 0.0
4:00 206.7 F 2.1
6:00 223.2 F 7.2
8:00 236.2 F 12.4
10:00 236.6 F 10.1
12:00 236.0 F 8.7
14:00 239.2 F 9.9
16:00 250.7 F 15.9
18:00 263.3 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.2 F 28.9
2:00 272.8 F 29.3
3:00 272.0 F 29.1
4:00 272.1 F 28.8

Min 271.6 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 223.0 F 0.0
10:00 231.0 F 0.0
15:00 227.3 F 0.0
20:00 223.7 F 0.0
25:00 221.1 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:60
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

08/15/2023 01:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____