

-----  
Midmark M11 - v1.0.0  
Total Cycles: 846

Name: RITTER 1  
Sterilizer ID: V2178198  
05/27/2021 02:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.8 F 0.0  
2:00 175.1 F 3.1  
4:00 189.9 F 7.2  
6:00 210.3 F 12.2  
8:00 218.3 F 6.6  
10:00 230.8 F 11.6  
12:00 233.2 F 8.4  
14:00 242.1 F 11.2  
16:00 245.1 F 12.8  
18:00 256.6 F 19.2  
20:00 267.6 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.6 F 29.1  
2:00 271.7 F 28.5  
3:00 271.2 F 28.5  
4:00 272.7 F 29.1  
  
Min 271.1 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 240.3 F 0.0  
10:00 219.7 F 0.0  
15:00 221.6 F 0.0  
20:00 205.3 F 0.0  
25:00 205.4 F 0.0  
30:00 215.9 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 21:25  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:20  
CYCLE COMPLETE

05/27/2021 03:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_