
Midmark M11 - v1.0.5
Total Cycles: 1775

Name: RITTER2
Sterilizer ID: V2313228
11/28/2022 10:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.8 F 0.0
2:00 192.5 F 2.3
4:00 191.6 F 7.2
6:00 214.2 F 12.8
8:00 225.6 F 8.4
10:00 237.6 F 11.0
12:00 242.5 F 12.8
14:00 242.6 F 12.2
16:00 256.3 F 19.4
18:00 268.3 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.8 F 30.0
2:00 271.9 F 29.3
3:00 271.4 F 29.5
4:00 272.9 F 30.0

Min 271.2 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 230.1 F 0.0
10:00 243.3 F 0.0
15:00 239.7 F 0.0
20:00 229.4 F 0.0
25:00 226.3 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

11/28/2022 11:29 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____