
Midmark M11 - v1.0.0
Total Cycles: 2535

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2022 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.1 F 0.0
2:00 183.3 F 2.8
4:00 184.1 F 7.1
6:00 202.1 F 12.0
8:00 219.8 F 6.4
10:00 232.0 F 12.0
12:00 233.6 F 8.8
14:00 239.3 F 9.7
16:00 246.7 F 13.5
18:00 259.1 F 20.3
20:00 270.4 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.1
2:00 271.5 F 28.6
3:00 272.5 F 29.1
4:00 271.5 F 28.5

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.1 F 0.0
5:00 228.9 F 0.0
10:00 240.5 F 0.0
15:00 232.5 F 0.0
20:00 231.1 F 0.0
25:00 223.3 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

11/09/2022 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____