
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000234
7:20 20/07/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:06
STERILIZING 4:05
136.0 C 220kPa 4:05
136.2 C 220kPa 4:35
135.9 C 220kPa 5:05
136.2 C 222kPa 5:35
136.4 C 221kPa 6:05
136.3 C 221kPa 6:35
136.4 C 219kPa 7:05
136.4 C 221kPa 7:35
136.4 C 221kPa 7:35
VENTING 7:36
AIR DRYING 8:25
CYCLE COMPLETE 9:12

Digital Signature #
5B92243CAEA77595
