
Midmark M11 - v1.0.0
Total Cycles: 4022

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2023 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.0 F 0.0
2:00 154.4 F 2.3
4:00 161.1 F 5.7
6:00 185.6 F 9.8
8:00 214.0 F 5.9
10:00 215.2 F 6.9
12:00 228.8 F 11.7
14:00 232.0 F 8.0
16:00 243.3 F 12.2
18:00 244.0 F 12.3
20:00 256.5 F 18.8
22:00 267.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.7 F 29.1
2:00 271.8 F 28.5
3:00 271.4 F 28.5
4:00 272.6 F 29.1

Min 271.1 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.1 F 0.0
5:00 222.7 F 0.0
10:00 231.2 F 0.0
15:00 228.3 F 0.0
20:00 229.3 F 0.0
25:00 228.8 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

11/15/2023 01:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____