

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002662  
16:08 07/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:04  
STERILIZING 3:12  
136.0 C 220kPa 3:12  
136.3 C 225kPa 3:42  
136.4 C 226kPa 4:12  
136.2 C 224kPa 4:42  
136.0 C 222kPa 5:12  
136.0 C 222kPa 5:42  
136.3 C 225kPa 6:12  
136.4 C 225kPa 6:42  
136.4 C 225kPa 6:42  
VENTING 6:43  
AIR DRYING 7:41  
CYCLE COMPLETE 12:15  
-----

Digital Signature #  
45AEDBAB19EE11F4  
-----