

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 008477  
11:46 27/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:58  
PRESSURIZING 2:02  
STERILIZING 3:01  
136.0 C 220kPa 3:01  
136.3 C 222kPa 3:31  
136.3 C 223kPa 4:01  
136.4 C 221kPa 4:31  
136.3 C 223kPa 5:01  
136.0 C 221kPa 5:31  
136.4 C 224kPa 6:01  
136.4 C 223kPa 6:31  
136.4 C 223kPa 6:31  
VENTING 6:31  
AIR DRYING 7:23  
CYCLE COMPLETE 21:00

-----  
Digital Signature #  
1C3379EB2836502B  
-----