
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001310
10:40 20/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:55
STERILIZING 4:00
136.4 C 223kPa 4:00
136.4 C 225kPa 4:31
136.2 C 223kPa 5:01
136.2 C 223kPa 5:31
136.4 C 225kPa 6:01
136.2 C 224kPa 6:31
136.4 C 225kPa 7:01
136.4 C 225kPa 7:31
136.4 C 225kPa 7:31
VENTING 7:31
AIR DRYING 8:22
CYCLE COMPLETE 23:22

Digital Signature #
9425BF46D48545A8
