
Midmark M11 - v1.0.0
Total Cycles: 1051

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2021 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 159.3 F 1.4
4:00 166.5 F 3.7
6:00 202.0 F 7.3
8:00 217.1 F 11.7
10:00 225.5 F 5.5
12:00 234.4 F 10.0
14:00 235.3 F 9.1
16:00 242.1 F 11.6
18:00 238.4 F 9.6
20:00 251.1 F 15.6
22:00 261.8 F 21.8
24:00 271.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.7 F 29.0
2:00 272.2 F 28.6
3:00 271.3 F 28.1
4:00 271.4 F 28.5

Min 271.0 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 229.4 F 0.0
10:00 184.9 F 0.0
15:00 186.5 F 0.0
20:00 171.5 F 0.0
25:00 177.5 F 0.0
30:00 180.6 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:13
CYCLE COMPLETE

07/26/2021 10:22 AM
ACCEPTED: X
REJECTED:
Operator: U01