
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008202
10:25 09/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:08
STERILIZING 3:09
136.4 C 221kPa 3:09
136.4 C 223kPa 3:39
136.0 C 221kPa 4:09
136.1 C 220kPa 4:39
136.4 C 223kPa 5:09
136.4 C 223kPa 5:39
136.2 C 223kPa 6:09
136.4 C 220kPa 6:39
136.4 C 220kPa 6:39
VENTING 6:39
AIR DRYING 7:26
CYCLE COMPLETE 14:52

Digital Signature #
3A75523830E9DCC9
