
Midmark M11 - v1.0.0
Total Cycles: 4160

Name: RITTER 1
Sterilizer ID: V2178198
12/27/2023 03:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.9 F 0.0
2:00 150.2 F 2.2
4:00 163.8 F 5.5
6:00 206.9 F 10.3
8:00 217.6 F 4.1
10:00 227.3 F 8.1
12:00 237.3 F 12.9
14:00 236.8 F 9.4
16:00 238.2 F 9.4
18:00 247.1 F 13.7
20:00 259.0 F 20.1
22:00 270.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.9 F 28.9
2:00 271.3 F 28.2
3:00 272.2 F 28.9
4:00 272.6 F 28.7

Min 271.0 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 249.0 F 0.0
10:00 228.7 F 0.0
15:00 226.1 F 0.0
20:00 225.5 F 0.0
25:00 227.0 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

12/27/2023 04:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____