
Midmark M11 - v1.0.5
Total Cycles: 2515

Name: RITTER2
Sterilizer ID: V2313228
06/14/2023 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.9 F 0.0
2:00 191.0 F 4.4
4:00 206.8 F 10.2
6:00 220.9 F 6.4
8:00 236.1 F 12.5
10:00 237.5 F 10.8
12:00 237.5 F 10.0
14:00 252.3 F 17.3
16:00 264.7 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.6 F 29.8
2:00 271.6 F 29.5
3:00 272.7 F 29.9
4:00 271.8 F 29.2

Min 271.5 F 29.2
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 231.7 F 0.0
10:00 229.1 F 0.0
15:00 231.3 F 0.0
20:00 225.8 F 0.0
25:00 228.2 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 17:47
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:21
CYCLE COMPLETE

06/14/2023 12:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____