
Midmark M11 - v1.0.0
Total Cycles: 5044

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2024 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 175.4 F 0.3
4:00 175.3 F 1.8
6:00 195.1 F 5.9
8:00 210.7 F 10.3
10:00 218.7 F 7.0
12:00 229.6 F 9.8
14:00 233.3 F 9.1
16:00 241.4 F 12.4
18:00 240.1 F 10.8
20:00 252.7 F 17.1
22:00 263.8 F 23.5
24:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 271.9 F 29.0
2:00 272.7 F 29.3
3:00 272.0 F 28.8
4:00 271.5 F 28.7

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 230.7 F 0.0
10:00 224.5 F 0.0
15:00 234.0 F 0.0
20:00 235.6 F 0.0
25:00 228.7 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

11/01/2024 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____