
Midmark M11 - v1.0.5
Total Cycles: 2832

Name: RITTER2
Sterilizer ID: V2313228
09/05/2023 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.7 F 0.0
2:00 165.1 F 3.2
4:00 188.6 F 7.9
6:00 208.0 F 13.2
8:00 226.2 F 8.6
10:00 234.7 F 10.9
12:00 241.0 F 12.3
14:00 241.1 F 11.4
16:00 254.6 F 18.3
18:00 266.2 F 25.4
20:00 273.1 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.5 F 29.7
2:00 272.1 F 29.8
3:00 272.1 F 29.6
4:00 271.9 F 29.8

Min 271.5 F 29.2
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 211.6 F 0.0
10:00 211.1 F 0.0
15:00 206.7 F 0.0
20:00 205.6 F 0.0
25:00 207.0 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

09/05/2023 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____