

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1038

Name: RITTER2  
Sterilizer ID: V2313228  
04/26/2022 03:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.6 F 0.0  
2:00 176.1 F 3.3  
4:00 195.6 F 7.9  
6:00 212.1 F 13.3  
8:00 224.5 F 8.1  
10:00 238.8 F 13.5  
12:00 239.7 F 11.5  
14:00 239.3 F 10.5  
16:00 253.2 F 17.5  
18:00 265.0 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.2 F 29.8  
2:00 272.6 F 29.6  
3:00 272.2 F 29.2  
4:00 271.2 F 28.8  
  
Min 271.0 F 28.8  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 231.5 F 0.0  
10:00 219.7 F 0.0  
15:00 227.3 F 0.0  
20:00 226.2 F 0.0  
25:00 226.6 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 19:54  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:09  
CYCLE COMPLETE

04/26/2022 04:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_