

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3827

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2023 11:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
2:00 183.4 F 2.6  
4:00 201.8 F 7.1  
6:00 216.5 F 13.0  
8:00 226.7 F 7.9  
10:00 241.4 F 13.3  
12:00 240.0 F 10.8  
14:00 238.0 F 9.2  
16:00 252.3 F 16.2  
18:00 264.3 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.4 F 28.5  
2:00 271.9 F 28.5  
3:00 272.3 F 28.4  
4:00 271.8 F 28.4  
  
Min 271.5 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 225.6 F 0.0  
10:00 235.1 F 0.0  
15:00 228.1 F 0.0  
20:00 227.0 F 0.0  
25:00 229.1 F 0.0  
30:00 228.7 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 19:54  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

09/27/2023 11:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_