
Midmark M11 - v1.0.0
Total Cycles: 1838

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2022 08:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.1 F 0.0
2:00 194.8 F 2.4
4:00 208.9 F 6.5
6:00 224.7 F 12.2
8:00 227.1 F 7.0
10:00 239.8 F 12.3
12:00 237.0 F 9.4
14:00 237.7 F 8.8
16:00 248.3 F 14.3
18:00 260.5 F 21.1
20:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 28.9
2:00 272.7 F 29.3
3:00 271.3 F 28.8
4:00 272.5 F 29.0

Min 271.3 F 28.6
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 232.8 F 0.0
10:00 234.2 F 0.0
15:00 232.1 F 0.0
20:00 213.9 F 0.0
25:00 225.4 F 0.0
30:00 232.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:20
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

04/19/2022 09:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____