
Midmark M11 - v1.0.5
Total Cycles: 2760

Name: RITTER2
Sterilizer ID: V2313228
08/14/2023 01:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 151.4 F 0.0
4:00 168.7 F 0.0
6:00 191.5 F 1.8
8:00 208.5 F 5.6
10:00 222.4 F 9.6
12:00 231.8 F 10.2
14:00 235.9 F 10.7
16:00 234.8 F 8.4
18:00 241.0 F 10.8
20:00 247.2 F 14.2
22:00 258.6 F 20.4
24:00 269.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.2
2:00 271.9 F 29.4
3:00 272.2 F 29.0
4:00 272.2 F 29.6

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 207.3 F 0.0
10:00 210.5 F 0.0
15:00 199.9 F 0.0
20:00 195.3 F 0.0
25:00 193.1 F 0.0
Dry Cycle Aborted

FILLING: 0:58
HEATING: 25:07
STERILIZING: 4:00
VENTING: 3:34
DRYING: 25:03
TOTAL CYCLE: 00:58:41

08/14/2023 02:32 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____