
Midmark M11 - v1.0.0
Total Cycles: 4603

Name: RITTER 1
Sterilizer ID: V2178198
06/03/2024 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.8 F 0.0
2:00 90.2 F 0.6
4:00 106.6 F 2.2
6:00 130.9 F 4.4
8:00 167.4 F 6.9
10:00 176.3 F 10.1
12:00 191.5 F 11.8
14:00 210.8 F 5.1
16:00 218.2 F 8.4
18:00 230.5 F 12.1
20:00 230.1 F 7.6
22:00 240.9 F 12.2
24:00 238.0 F 9.7
26:00 249.4 F 15.1
28:00 259.3 F 20.6
30:00 268.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.0
2:00 271.2 F 28.3
3:00 272.3 F 29.2
4:00 271.7 F 28.6

Min 271.1 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 213.6 F 0.0
10:00 211.5 F 0.0
15:00 206.5 F 0.0
20:00 207.0 F 0.0
25:00 208.6 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:15
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

06/03/2024 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____