

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001034  
14:52 08/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:44  
STERILIZING 3:44  
136.0 C 220kPa 3:44  
136.4 C 225kPa 4:14  
136.4 C 223kPa 4:44  
136.1 C 223kPa 5:14  
136.3 C 225kPa 5:44  
136.4 C 223kPa 6:14  
136.1 C 223kPa 6:44  
136.4 C 225kPa 7:14  
136.4 C 225kPa 7:14  
VENTING 7:14  
AIR DRYING 8:05  
CYCLE COMPLETE 23:05  
-----

Digital Signature #  
B5EE41CEEC09C1AA  
-----