
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003607
8:05 05/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:12
STERILIZING 4:20
136.5 C 223kPa 4:20
136.1 C 221kPa 4:50
136.4 C 223kPa 5:20
136.1 C 219kPa 5:50
136.4 C 224kPa 6:20
136.3 C 224kPa 6:50
136.4 C 220kPa 7:20
136.2 C 218kPa 7:50
136.2 C 218kPa 7:50
VENTING 7:50
AIR DRYING 8:37
CYCLE COMPLETE 9:47

Digital Signature #
4CBF6BA0E694F3CB
