
Midmark M11 - v1.0.5
Total Cycles: 2364

Name: RITTER2
Sterilizer ID: V2313228
05/15/2023 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.3 F 0.0
2:00 134.3 F 0.0
4:00 142.4 F 0.1
6:00 176.9 F 3.0
8:00 198.5 F 7.4
10:00 215.8 F 12.5
12:00 223.2 F 7.3
14:00 235.1 F 12.3
16:00 236.4 F 9.8
18:00 237.3 F 9.2
20:00 248.4 F 14.7
22:00 260.0 F 21.2
24:00 270.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.5
2:00 271.6 F 28.7
3:00 272.7 F 29.6
4:00 272.1 F 29.0

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 202.1 F 0.0
10:00 215.4 F 0.0
15:00 211.7 F 0.0
20:00 223.9 F 0.0
25:00 218.6 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:50
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:53
CYCLE COMPLETE

05/15/2023 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____