

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3938

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2024 02:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.2 F 0.0  
2:00 146.1 F 2.1  
4:00 153.0 F 5.1  
6:00 178.9 F 9.4  
8:00 200.5 F 9.5  
10:00 216.1 F 7.6  
12:00 228.7 F 12.0  
14:00 231.9 F 8.7  
16:00 239.9 F 11.5  
18:00 243.9 F 12.7  
20:00 254.9 F 18.8  
22:00 265.9 F 25.3  
24:00 272.1 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.6 F 29.8  
1:00 272.7 F 30.0  
2:00 272.7 F 30.2  
3:00 272.4 F 30.3  
4:00 271.7 F 29.9

Min 271.7 F 29.7  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
5:00 231.7 F 0.0  
10:00 215.0 F 0.0  
15:00 209.9 F 0.0  
20:00 206.1 F 0.0  
25:00 202.1 F 0.0  
30:00 212.7 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 24:20  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:03:05  
CYCLE COMPLETE

09/12/2024 03:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_