

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 004292  
8:26 06/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:09  
STERILIZING 4:13  
136.4 C 222kPa 4:13  
135.8 C 221kPa 4:43  
136.2 C 220kPa 5:13  
136.4 C 223kPa 5:43  
136.2 C 219kPa 6:13  
136.3 C 222kPa 6:43  
136.3 C 220kPa 7:13  
136.2 C 223kPa 7:43  
136.2 C 223kPa 7:43  
VENTING 7:43  
AIR DRYING 8:31  
CYCLE COMPLETE 10:47

-----  
Digital Signature #  
B148D81747721B3A  
-----