
Midmark M11 - v1.0.5
Total Cycles: 3127

Name: RITTER2
Sterilizer ID: V2313228
11/15/2023 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 169.2 F 2.4
4:00 178.2 F 6.4
6:00 201.7 F 11.1
8:00 216.0 F 5.3
10:00 224.6 F 10.8
12:00 230.1 F 7.8
14:00 239.2 F 11.9
16:00 244.1 F 13.2
18:00 256.8 F 19.8
20:00 268.2 F 26.9
22:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 272.8 F 30.1
2:00 271.9 F 30.1
3:00 272.4 F 29.9
4:00 272.1 F 30.3

Min 271.7 F 29.6
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 226.0 F 0.0
10:00 247.5 F 0.0
15:00 242.3 F 0.0
20:00 231.6 F 0.0
25:00 224.2 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

11/15/2023 01:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____