
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003382
14:00 04/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:00
STERILIZING 4:02
136.0 C 220kPa 4:02
136.4 C 223kPa 4:32
136.2 C 224kPa 5:02
136.4 C 221kPa 5:32
136.4 C 221kPa 6:02
136.1 C 223kPa 6:32
136.1 C 222kPa 7:02
136.4 C 223kPa 7:32
136.4 C 223kPa 7:32
VENTING 7:32
AIR DRYING 8:29
CYCLE COMPLETE 8:32

Digital Signature #
249958D70F6AB0E5
