

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3096

Name: RITTER 1  
Sterilizer ID: V2178198  
04/12/2023 07:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.8 F 0.0  
2:00 81.5 F 0.0  
4:00 93.8 F 0.0  
6:00 129.7 F 0.2  
8:00 153.0 F 2.1  
10:00 179.9 F 4.9  
12:00 195.2 F 8.4  
14:00 211.4 F 12.1  
16:00 218.6 F 5.4  
18:00 229.6 F 9.6  
20:00 236.8 F 10.3  
22:00 240.7 F 11.2  
24:00 237.3 F 9.3  
26:00 251.0 F 15.8  
28:00 262.4 F 22.3  
30:00 272.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.8  
2:00 271.6 F 28.4  
3:00 272.4 F 28.8  
4:00 271.5 F 28.3

Min 271.5 F 28.2  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.7 F 0.0  
5:00 199.0 F 0.0  
10:00 207.3 F 0.0  
15:00 226.9 F 0.0  
20:00 218.0 F 0.0  
25:00 216.9 F 0.0  
30:00 217.0 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 30:20  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:08:19  
CYCLE COMPLETE

04/12/2023 08:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_