
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000152
14:29 14/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:14
PRESSURIZING 4:14
STERILIZING 5:36
136.4 C 225kPa 5:36
136.4 C 226kPa 6:06
136.4 C 226kPa 6:36
136.3 C 224kPa 7:06
136.3 C 224kPa 7:36
136.1 C 225kPa 8:06
136.4 C 225kPa 8:36
136.3 C 225kPa 9:06
136.3 C 225kPa 9:06
VENTING 9:06
AIR DRYING 10:08
CYCLE COMPLETE 15:22

Digital Signature #
98F3D48B0630AB07
