

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3038

Name: RITTER2  
Sterilizer ID: V2313228  
10/20/2023 02:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.8 F 0.0  
2:00 152.5 F 2.4  
4:00 164.8 F 5.8  
6:00 191.8 F 10.3  
8:00 210.2 F 5.0  
10:00 227.6 F 9.4  
12:00 232.6 F 9.4  
14:00 240.3 F 12.1  
16:00 239.7 F 10.9  
18:00 252.8 F 17.5  
20:00 264.4 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.5 F 29.7  
2:00 271.8 F 29.7  
3:00 272.5 F 29.7  
4:00 271.6 F 29.5  
  
Min 271.5 F 29.2  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 233.8 F 0.0  
10:00 218.3 F 0.0  
15:00 214.6 F 0.0  
20:00 209.4 F 0.0  
25:00 213.4 F 0.0  
30:00 217.7 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 21:56  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:00:26  
CYCLE COMPLETE

10/20/2023 03:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_